

**SOUTH CAROLINA LAW ENFORCEMENT DIVISION**

4400 BROAD RIVER ROAD  
COLUMBIA, SOUTH CAROLINA 29210

POLICY #: 8.12.19	DATE: 09/01/93	REVISION DATE: 11/06/98
TITLE: IMPLIED CONSENT - BAC DATAMASTER SIMULATOR SOLUTION TEMPERATURE		PAGE 1 OF 1

RESPONSIBLE AUTHORITY: SPECIAL AGENT IN CHARGE OF  
FORENSIC SERVICES LABORATORY


RELATED STANDARDS/STATUTES/REFERENCES: S.C. Code of Laws, Section 56-5-2950  
S.C. Code of Laws, Section 50-21-114  
S.C. Code of Laws, Section 55-1-100  
S.C. Code of Laws, Section 56-1-2130  
S.C. Code of Laws, Section 23-31-410  
S.C. Code of Laws, Section 56-1-286

**GENERAL PURPOSE:** To set forth policies for the administration of implied consent breath alcohol tests.

**POLICY:** The Division will establish procedures for properly conducting implied consent BAC DataMaster breath alcohol tests.

**SPECIFIC PROCEDURES:** During a subject test, the simulator solution temperature is checked by the probe at the beginning of the test sequence, and also at the time of the simulator test. Depending on the software version, the simulator temperature may also be checked during the time stamp procedure. The simulator solution temperature must be 34.0 degrees Celsius (C) plus or minus 0.5 degrees C, to pass each test. If the simulator solution temperature is in this range, 34c will be printed on the evidence ticket. If the simulator solution temperature is not in this range, the test is aborted.

BY ORDER OF:

  
\_\_\_\_\_  
CHIEF OF SLED